

Test Report

No.: AJHL1607007461FT

Date: JUL.18, 2016

Page 1 of 3

NINGBO TRANSO-MEDICAL CO., LTD.

ROOM1608, NO. 18 DIAMOND COMMERCIAL PLAZA, NO. 409 JIANG'AN ROAD, JIANGBEI DISTRICT, NINGBO 315021, CHINA

The following sample(s) was/were submitted and identified on behalf of the applicant as:

Product Description	CARRYING COT
SGS Ref No.	NBHL1607009703SD
Sample Receiving Date	JUL.13, 2016
Test Performing Date	JUL.13, 2016 TO JUL.15, 2016

Test Result Summary	
Test(s) Requested	Result(s)
EN 1865-1:2010+A1:2015, Clause 4.5.1, 4.5.2, 4.5.3, 4.5.4, 4.5.8, 4.5.10	PASS
Summary:	
1. For further details, please refer to the following page(s)	

Signed for and on behalf of
SGS-CSTC Co., Ltd. Anji BranchDavid Fan
Approved SignatorySGS-CSTC Standards Technical Services Co., Ltd.
Anji Branch Standard Testing Services Laboratory

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No. 301, Sunlight Road, 2 Block, Sunlight Industry Zone, Anji County, Zhejiang Province, China 313300 t (86-572) 5018825 f (86-572) 5018829 www.sgsgroup.com.cn
中国·浙江·安吉县阳光工业园二区阳光大道301号 邮编:313300 t (86-572) 5018825 f (86-572) 5018829 e sgs.china@sgs.com

Test Report

No.: AJHL1607007461FT

Date: JUL.18, 2016

Page 2 of 3

Test Conducted: EN 1865-1:2010+A1:2015 Patient handling equipment used in road ambulances
Part 1: General stretcher systems and patient handling equipment

Test Result:

Test Item	Test Method & Test Requirement	Test Result
4.5 Carrying sheet		
4.5.1 Dimensions	The dimensions of the carrying sheet shall be as follows: — Length: minimum 1850mm; — Width: minimum 570mm.	Pass
4.5.2 Mass	The mass shall be not more than 5 kg.	Pass
4.5.3 Loading capacity	The loading capacity shall be a minimum of 150 kg.	Pass
4.5.4 Handles	The carrying sheet shall be equipped with at least three handles on each longitudinal side.	Pass
4.5.8 Deformation of the handles	There shall be no remaining deformation of the handles. Test in accordance with 5.4.1: a) Apply a force of $F = 2500$ N between all opposite carrying handles. b) Apply a force of $F = 1000$ N between the outer and middle handles of each side. There shall be no broken threads upon completion of the above tests.	Pass
4.5.10 Deformation of the lying area	There shall be no remaining deformation of the lying area. Test in accordance with 5.4.2: Secure carrying sheet handles, along the length of the carrying sheet, to support. Carrying sheet will be suspended from the supports. Load carrying sheet with 250 kg, distributing the weight evenly along the length of the carrying sheet. Unload sheet and examine for failures.	Pass

Remark:

1. Sample overall dimension: 1880mm (L)×750mm (W); Sample weight: 1.35kg



This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No. 301, Sunlight Road, 2 Block, Sunlight Industry Zone, Anji County, Zhejiang Province, China 313300 t (86-572) 5018825 f (86-572) 5018829 www.sgsgroup.com.cn
 中国·浙江·安吉县阳光工业园二区阳光大道301号 邮编:313300 t (86-572) 5018825 f (86-572) 5018829 e sgs.china@sgs.com

Test Report

No.: AJHL1607007461FT

Date: JUL.18, 2016

Page 3 of 3

Photo Appendix:



SGS authenticate the photo on original report only

End of Report



This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /Inspection report & certificate, please contact us at telephone: (86-572) 5018829, or email: CN_Doccheck@sgs.com

No. 301, Sunlight Road, 2 Block, Sunlight Industry Zone, Anji County, Zhejiang Province, China 313300 t (86-572) 5018829 f (86-572) 5018829 www.sgsgroup.com.cn
中国·浙江·安吉县阳光工业园二区阳光大道301号 邮编:313300 t (86-572) 5018829 f (86-572) 5018829 e sgs.china@sgs.com